



CERTIFICATE OF ANALYSIS

Solstice Advanced Materials

1953 South Harvey Street

Muskegon, MI 49442

Phone: (800) 368-0050

lab.solstice.com**Brand:**

CHROMASOLV™

Product:

39292

Lot No.:

Example

Date of Manufacture:

3-Aug-2026

Recommended Retest Date:

1-Feb-2028

Water, LC-MS for PFAS Analysis

Parameter	Specification		Result	Units
	Min.	Max.		
Residue		1	<0.5	mg/L
Ultraviolet Absorbance Test	Passes test		Passes test	
Refractive Index (20°C)	1.3325	1.3335	1.3332	
HPLC Gradient Elution	Passes test		Passes test	
Fluorescence (as quinine base)		0.30	0.20	ppb
LC-MS Positive Mode (as reserpine)		50	<50	ppb
LC-MS Positive Mode (as caffeine)		100	<100	ppb
LC-MS Negative Mode (as 4-nitrophenol)		50	<50	ppb
Trace Metals	Passes test		Passes test	
Perfluoro-n-butanoic Acid (PFBA)		10.0	<10.0	ppt
Perfluorobutanesulfonic Acid (PFBS)		5.0	<5.0	ppt
Perfluorohexanesulfonic Acid (PFHxS)		5.0	<5.0	ppt
Perfluoro-n-octanoic Acid (PFOA)		5.0	<5.0	ppt
Perfluorooctanesulfonic Acid (PFOS)		5.0	<5.0	ppt

Solstice Advanced Materials

Quality Control Approval

Muskegon

8/19/2026

LIMS Sample No.:

AR01942